



Cat. No.	Product Description
CE-SP023	<p>Equipment is used for testing Hydraulic Cement under Wet Washing Condition as per IS: 4031 (Part-15).</p> <p>Equipment consists of Inlet Valve with Nozzle, Pressure Gauge of 100mm dia x 2kg/cm² (Applied Water Pressure is 0.7138 kg/cm² or 10.15 PSI with 1gm of sample) with NABL calibration certificate and Spray Nozzle. Complete unit is made from non corrosive material. Unit is supplied with Brass Frame Sieve 50mm dia or 100mm dia x 45microns wire mesh.</p>